



California Institute of Technology
Massachusetts Institute of Technology

DCN No. E020525-00-C

Document Change Notice (DCN)

Sheet 1 of 1

DOCUMENT No. (DOC-REV-GP.ID)	TITLE	NEW REV
D980325-B	Optical Lever Photodiode Amplifier	C

CHANGE DESCRIPTION (FROM/TO):

The schematic and circuit board have been modified to accept either QD100 or QD50 photodiodes. All other board functions remain the same.

REASON FOR CHANGE: QD50 photodiodes are needed for some applications.

ACTION: Incorporate Change Attach DCN to Drawings Other Action (specify):

DISPOSITION OF HARDWARE (IDENTIFY SERIAL NUMBERS)	DCN DISTRIBUTION															
<input type="checkbox"/> No hardware was affected (record change only):	<table> <tr><td>Barish</td><td>Coles</td><td>Coyne</td></tr> <tr><td>Lazzarini</td><td>Lindquist</td><td>Sanders</td></tr> <tr><td>Shoemaker</td><td>Stapfer</td><td>Tyler</td></tr> <tr><td>Weiss</td><td>Whitcomb</td><td>Matherny</td></tr> <tr><td>Raab</td><td>McCarthy</td><td>Wooley</td></tr> </table>	Barish	Coles	Coyne	Lazzarini	Lindquist	Sanders	Shoemaker	Stapfer	Tyler	Weiss	Whitcomb	Matherny	Raab	McCarthy	Wooley
Barish		Coles	Coyne													
Lazzarini		Lindquist	Sanders													
Shoemaker		Stapfer	Tyler													
Weiss		Whitcomb	Matherny													
Raab		McCarthy	Wooley													
<input type="checkbox"/> List S/Ns which comply already:																
<input type="checkbox"/> List S/Ns to be reworked/scrapped:																
<input type="checkbox"/> List S/N's to be built with this change:																
<input type="checkbox"/> List S/Ns to be retested per this change:																
<input checked="" type="checkbox"/> All new boards produced will be rev C.																

SAFETY, COST, SCHEDULE, REQUIREMENTS IMPACT? NO YES (If YES, enter CR (CCB) or TCP (TRB) #)

APPROVALS:	DATE	OTHER APPROVALS (SPECIFY)	DATE
ORIGINATOR:	5/2/02		
TASK LEADER:			
GROUP LEADER:			
DCC RELEASE:			