*LIGO*-E980130- *B*- *D*DRWG NO. REV.
GID

SHEET 1 OF

### **COMPONENT SPECIFICATION**

# TEL PAR SECONDARY MIRROR

APPROVALS:	DATE	REV	DCN NO	BY	СНК	DCC	DATE
DRAWN: MICHAEL SMITH, JONATHAN KERN	10/6/98	Α	E980325-00-D				11/5/98
CHECKED:		В	E990079-00-D				1/20/99
APPROVED:							
DCC RELEASE:							

## 1 SCOPE

This is a specification for the TEL\_PAR Secondary Mirror. This convex parabolic secondary mirror, together with the concave parabolic primary mirror, forms an off-axis Mersenne reflecting 8X beam-reducing telescope.

### 2 APPLICABLE DOCUMENTS

- 1) Ligonian VI-B, 8x magnification, 1.3m, LIGO-D980xxx-00-D
- 2) Pick Off Telescope, LIGO-E980262-03-D
- 3) MIL-C-675C

# 3 REQUIREMENTS

### 3.1. PERFORMANCE CHARACTERISTICS

#### 3.1.1. Mirror Fabrication

#### 3.1.1.1 Front Surface

Radius of curvature -381 mm + /-4Conic constant -1.00 + /-0.02Parabola tilt, reference to back surface +/-0.05 degClear aperture -381 mm + /-4 + /-0.05 deg-1.00 + /-0.05 deg

Off-axis displacement 26.20 + -0.5 mmDiameter 34.0 + 0.0, -0.1 mmMinimum edge thickness 20.0 + -0.6 mm

Surface irregularity <0.25 wave peak to valley@ 633 nm over clear aperture

Surface finish 60/40 Surface roughness <100 Ang

#### 3.1.1.2 Back Surface

Surface quality fine ground, >300 grit

*LIGO*-E980130-B-DDRWG NO. REV. GID

SHEET 2 OF **CONTINUATION SHEET** 

## **COMPONENT SPECIFICATION**

# TEL PAR SECONDARY MIRROR

#### 3.1.1.3 **Mirror Coating 1**

Apply to primary and secondary mirror front surfaces

Front surface only

Wavelength 1064 nm

Polarization Incidence angle 4 deg Multilayer dielectric coating, Reflectivity >99.9%

Durability per MIL-C-675C

#### 3.1.1.4 **Mirror Coating 2**

Apply to primary and secondary mirror front surfaces

Front surface only

Wavelength 1064 nm

Polarization S Incidence angle 4 deg Protected Silver, Reflectivity >98%

Durability per MIL-C-675C

LIGO-E980130- B- D

DRWG NO. REV.
GID

SHEET 3 OF 7

**CONTINUATION SHEET** 

# **COMPONENT SPECIFICATION**

# TEL\_PAR SECONDARY MIRROR

12:56:29 D980263-00-D В RADIUS RAD TOL łR C.A. DIA EDGE EDGE DIA DIA TOL THICKNESS THI TOL WEDGE -381.0 +4.0 32.0 .25 34.0 20.0 +/-0.6 +0.0, -0.1 S2 NOTES: Etch Serial 1. All Dimensions are in mm Number 2. Material: Zerodur or fused Silica 3. K=-1.00 + / - 0.02Top View 4. Min Edge: 20.0 + / - 0.65. Center offset: 26.20 + /-0.56. Parabola tilt: +/- 0.05 deg ref back surface Tic-Mark 7. Surface Finish: 60/40 8. Surface roughness:<100 Ang 9. Chamfer edges: 0.25 mm, 45 + / - 10 deg10. Back Surface: fine ground > 300 grit Optic Axis 11. Edge tic-mark on edge nearest optic axis to indicate orientation of mirror. 0.24 +/- 0.05 wide x 15 +/- 2 long Side View 12. S1 Coating: > 99.9% reflectance @ 1.064 um Durability: per MIL-C-675C 20.0 Polarization: (s) OFF-AXIS PARABOLA Incidence Angle: 4 deg Clif. Institute of Technology D. Rose 13. Etch serial number lettering 06/15/98 LIGO Project M. Smith 06/15/98 approximately 4 mm high 1200 E. California Blvd. Pasadena, CA, 91125 LigonianVIB, Secondary D980263-00-D В file:Ligo8xse.dwg

Figure 1: Fabrication drawing, PO telescope secondary mirror

### 3.2. DESIGN AND CONSTRUCTION

#### 3.2.1. Materials

#### **3.2.1.1** Mirrors

Substrate material Zerodur or fused silica

LIGO-E980130- B- DDRWG NO. REV.
GID

SHEET 4 OF 7
CONTINUATION SHEET

# COMPONENT SPECIFICATION

# TEL\_PAR SECONDARY MIRROR

Edge chamfer

0.25 mm @ 45 deg +/- 10deg

## 3.3. Quality Assurance/Control

#### 3.3.1. Identification

Separate (non-welded) parts and assemblies shall be marked with laser marking or acid etch techniques. A vibratory tool with a minimum tip radius of 0.0005" is acceptable for marking on surfaces which are not hidden from view. Engraving is also permitted.

Separate (non-welded) parts and assemblies to be serialized according to the document titled Part Numbers and Serialization of Detector Hardware, LIGO-L970196. This document allows for "bag-and-tag" type of identification for small parts.

#### 3.3.1.1 Optical Alignment Marks

Edge tick-mark shall be placed to indicate orientation of mirror next to the optical axis:

0.24 + -0.05 mm wide x 20 mm + -2 mm long.

#### 3.3.1.2 Serial Number

#### 3.3.1.2.1 Optical Serial Number

A serial number identifying a component set of primary and secondary mirrors shall be etched, ground or sandblasted next to the alignment mark.

#### 3.3.1.2.1.1 Serial Number Format

The Serial number shall be of the format:

E980130-y S/N nnn-c Where

E980130-y is the LIGO piece part or assembly drawing number, E980130, including the revision letter, -y, to which the hardware item was built;

nnn is the sequential serial number, 001 through 999, in the order produced,

and c=1 describes the dielectric mirror coating, c=2 describes the protected silver mirror coating.

### **3.3.2.** Quality Assurance Provisions

A first article shall be produced and inspected for form, dimensions and workmanship.

#### 3.3.3. Purchaser Access

Non-escort privileges for the buyer, owner, government and owner representatives to all areas of the facilities where work is being performed shall be arranged. This will include access to all areas where material is being processed and stored. The purchaser shall have the right to witness all manufacturing processes.

*LIGO*-E980130- *B- D*DRWG NO. REV.
GID

SHEET 5 OF 7
CONTINUATION SHEET

# COMPONENT SPECIFICATION

# TEL\_PAR SECONDARY MIRROR

### 3.3.4. QA Approval

LIGO QA reserves the right to inspect and approve vendor/fabricator QA plan and processes.

### 4 TEST PROCEDURES

The telescope mirror performance shall be verified by the following test procedures

### 4.1. WAVEFRONT DISTORTION

Interferogram of transmitted wavefront across clear aperture, @1064nm wavelength

### 5 DOCUMENTATION

- 1) Compliance Certification for items 3.1., 3.2., 3.3., 4.
- 2) Interferogram, See "WAVEFRONT DISTORTION" on page 5.

### 6 ENVIRONMENTAL CHARACTERISTICS

The PO Telescope will operate in a non-vibrational, ultra high vacuum environment, at room temperature (68F,+/-4F).

# 7 HANDLING AND SHIPPING PROCEDURES

## 7.1. Cleaning

Approved cleaning procedures for UHV components are detailed in LIGO-E960022, Vacuum Compatibility, Cleaning Methods and Compatibility Procedures.

## 7.1.1. Optical Surfaces

All optical surfaces shall be cleaned in accordance with good commercial practice. Nothing shall contact the optical surfaces after cleaning, except for lint-free lens tissue.

# 7.2. Packaging for Shipment

# 7.2.1. Optical Parts

The cleaned optical components shall be protected with lint-free lens tissue. In addition, all components shall be placed in a sealed, clean polyethylene bag before shipping.

 $\begin{array}{ccc} LIGO\text{-}E980130\text{-} & B\text{-} & D \\ & \text{DRWG NO.} & \text{REV.} \\ \text{GID} & & & \end{array}$ 

SHEET 6 OF 7
CONTINUATION SHEET

# **COMPONENT SPECIFICATION**

# TEL\_PAR SECONDARY MIRROR

The shipping containers tected during transit.	must ensure	that the bag	does not get	punctured and the	hat the parts are	properly pro-



 $\begin{array}{ccc} LIGO\text{-}E980130\text{-} & B\text{-} & D \\ & \text{DRWG NO.} & \text{REV.} \\ \text{GID} & & & \end{array}$ 

COMPONENT SPECIFICATION	SHEET 7 OF 7 CONTINUATION SHEET
TEL_PAR SECONDARY MIRROR	