

Item # 2



375 Saxonburg Boulevard
Saxonburg, PA 16056
U.S.A

Coating Type	AOI	Pol.
S1 AR	0°	R
S2 AR	0°	R

Wavelength/s 10.6 μm

% Transmission % Reflection

Traced On:

Side 1 Side 2 Side 1 and Side 2

Coating Run No. 150-157866

Notes

Date 6/27/2011

Witness Sample ZnSe

Scale 0-100%

Operator RPS

Spectrophotometer traces are appropriate for determining the centering and reflectivity or transmission performance of a coating at a particular wavelength range. The reflectivity or transmission shown at any given wavelength may be $\pm 3\%$ of the actual value. The most accurate method of determining the reflectivity or transmission of a coating at its design wavelength is by laser test. Contact your sales representative for more information on laser tests or test data for specific optics.

