
NOTES CONTINUED: (5) SCRIBE, ENGRAVE (A VIBRATORY TOOL MAY BE USED), LASER MARK OR MECHANICALLY STAMP (NO INKS OR DYES) DRAWING PART NUMBER, REVISION (AND VARIANT OR "TYPE" IF APPLICABLE) ON NOTED SURFACE OF PART FOLLOWED ON THE NEXT LINE WITH A THREE DIGIT FOR THE FIRST. SERIAL NUMBERS START AT 001 FOR THE FIRST ARTICLE AND PROCEED CONSECUTIVELY, USE MINIMUM 0.12" HIGH CHARACTERS, UNLESS THE SIZE OF THE PART DICTATES SMALLER CHARACTERS. EXAMPLE: DXXXXXXX-VY, TYPE-XX, S/N XXX 6. APPROXIMATE WEIGHT: 2.97 LB 7. MACHINE ALL SURFACES TO REMOVE OXIDES AND MILL FINISH, USE OF ABRASIVE REMOVAL TECHNIQUES IS NOT ALLOWED. REFER TO LIGO-E0900364 8. ALL PARTS SHALL BE MANUFACTURED IN ACCORDANCE WITH LIGO SPECIFICATION E0900364. 9. ALL MATERIAL IS TO BE VIRGIN MATERIAL (i.e. NO WELD REPAIRS, PLUGS OR RECYCLED MATERIAL). NO REPAIRS SHALL BE MADE UNLESS APPROVED IN ADVANCE, AND IN WRITING, BY LIGO LABORATORY. REFER TO LIGO-E0900364. 22.5° 45.0° 8X ∅.281 THRU ALL EQ. SP. ON A Ø 5.000 B.C $| \oplus |$.014 $| \land A |$ B

.005 M A


R .25




REV.	DATE	DCN#	DRAWING TREE #
v 1	13 APR 2010	E1100356-x0	-
-	-	-	-
-	-	-	-


2

3



DETAIL A SCALE 1 : 1





NOTES AND TOLERANCES: (UNLESS OTHERWISE SPECIFIED)			ZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZZ	TECHNOLOGY	PART NAME				
A DIMENSIONS ARE IN INCHES	1. INTERPRET DRAWING PER ASME Y14.5-1994. 2. REMOVE ALL SHARP EDGES, .005015. FOR MACHINED PARTS. ROUND		LIGO CALIFORNIA INSTITUTE OF TECHNOLOGY MASSACHUSETTS INSTITUTE OF TECHNOLOGY		ALIGO, OPLEV, PERISCOPE, MT. TUBE				
TOLERANCES: ALL EDGES APPROXIMATLEY R.02 FOR SHEET METAL PARTS. 3. DO NOT SCALE FROM DRAWING.			SUB-SYSTEM	DESIGNER	E.JAMES	21 FEB 2011 SIZE	DWG. NO.	REV.	
.XX ± .01 .XXX ± .005	4. ALL MACHINING FLUIDS MUST BE FULLY SYNTHETIC, FULLY WATER SOLUBLE AND FREE OF SULFUR, SILICONE, AND CHLORINE.		ADVANCED LIGO (OPLEV	DRAFTER	E.SANCHEZ	08 APR 2011	c D1100898	
	MATERIAL	FINISH	NEXT ASSY	·	CHECKER	J.LEWIS	10 MAY 2011	D1100070	V 1
ANGULAR ± 1.0°	6061-T6	63 µin	ch D1100625		APPROVAL		SCA	ALE: 1:2 PROJECTION:	SHEET 1 OF 1