NOTES CONTINUED:

(5) SCRIBE, ENGRAVE, OR MECHANICALLY STAMP (NO INKS OR DYES) DRAWING PART NUMBER AND REVISION ON NOTÉD SURFACE FOLLOWED ON THE NEXT LINE BY A THREE DIGIT SERIAL NUMBER, SERIAL NUMBERS START AT 001 FOR THE FIRST ARTICLE AND PROCEED CONSECUTIVELY. USE .07" HIGH CHARACTERS. DXXXXXXX-VY, S/N 001.

VIBRATORY TOOL MAY BE USED. 6. APPROXIMATE WEIGHT = X.XXX LB.

7. MACHINE ALL SURFACES TO REMOVE OXIDES AND MILL FINISH, USE OF ABRASIVE REMOVAL TECHNIQUES IS NOT ALLOWED. REFER TO LIGO-E0900364

10. ALL HELI-COIL INSERTS TO BE INSTALLED BY LIGO PERSONNEL, AFTER DELIVERY OF FINISHED PARTS, USE NITRONIC 60 THREADED INSERTS.

11. ALL MATERIAL IS TO BE VIRGIN MATERIAL (i.e. NO WELD REPAIRS, PLUGS OR RECYCLED MATERIAL). NO REPAIRS SHALL BE MADE UNLESS APPROVED IN ADVANCE, AND IN WRITING, BY LIGO LABORATORY.

12. SURFACE FINISH TO BE AS-PROCESSED FROM MILL/SUPPLIER, FREE FROM SCRATCHES OR GOUGES.


13. PART WILL BE PORCELAIN COATED PER LIGO SPECIFICATION E1000083
AFTER FABRICATION. THE INDICATED HOLES WILL BE MASKED PRIOR TO
PORCELAIN COATING TO APPROXIMATELY 2.5-3X HOLE DIAMETER

14. DIMENSIONS APPLY BEFORE PORCELAIN COATING UNLESS SPECIFIED.

15. BEND RADIUS: UNLESS OTHERWISE NOTED, THE BEND RADIUS SHOULD BE THE MINIMUM REQUIRED TO FORM WITHOUT CRACKING OR REQUIRING ADDITIONAL WORK WHEN FORMING. IN PARTICULAR IF SHEET METAL IS TO BE PORCELAIN COATED THE BEND RADIUS SHALL BE A MINIMUM OF .12" OUTSIDE RADIUS OF BEND UNLESS


NOTES 9, 10, 13 and 14 DO NOT APPLY TO THIS PART

DATE DCN# DRAWING TREE #




V25A-108 CABLE ASSEMBLY CIRCUIT SUMMARY V-DB25 F/S1-108-DB25 F/S1						
CABLE NAME	COND WIRE ID	TWISTED PAIR	LENGTH	FROM	то	
V25A-108	25 COND. CABLE	(12 TOTAL)	108 in.	Conn. J1	Conn. J	
	W1	SHIELD	108 in	PIN 1, SHELL	PIN 1, SHELL	
	W2 TD 1	108 in	PIN 2	PIN 2		
	W14	TP-1	108 in	PIN 14	PIN 14	
	W3	TD O	108 in	PIN 3	PIN 3	
	W15	TP-2	108 in	PIN 15	PIN 15	
	W4	W4 TP_3	108 in	PIN 4	PIN 4	
	W16		108 in	PIN 16	PIN 1	
	W5	TD 4	108 in	PIN 5	PIN 5	
	W17	TP-4	108 in	PIN 17	PIN 17	
	W6 TD /	TD <i>E</i>	108 in	PIN 6	PIN 6	
	W18	TP-5	108 in	PIN 18	PIN 18	
	W7	TP-6	108 in	PIN 7	PIN 7	
	W19		108 in	PIN 19	PIN 1	
	W8	TP-7	108 in	PIN 8	PIN 8	
	W20	117-7	108 in	PIN 20	PIN 2	
	W9	TP-8	108 in	PIN 9	PIN 9	
	W21	11 -0	108 in	PIN 21	PIN 2	
	W10	TP-9	108 in	PIN 10	PIN 1	
W22	11 - 7	108 in	PIN 22	PIN 2		
	W11	TP-10	108 in	PIN 11	PIN 1	
W23 W12 W24 W13 W25 TP-10 TP-10	W23	11 -10	108 in	PIN 23	PIN 2	
	W12	TD 11	108 in	PIN 12	PIN 1	
	W24	15-11	108 in	PIN 24	PIN 2	
	W13	TD 10	108 in	PIN 13	PIN 1	
	7 17-12	108 in	PIN 25	PIN 2		

W25 * THE LENGTH SHOWN IN THIS LIST IS THE LENGTH OF THE CABLE BETWEEN THE TWO CONNECTORS. ADD ADDITIONAL LENGTH AS NECESSARY FOR THE INTERNAL WIRING OF THE CONNECTORS AND STRIP LENGTH.



NOTES AND TOLERANCES: (UNLESS OTHERWISE SPECIFIED)		JIIII CALIFORNIA	CALIFORNIA INICTITUTE OF TECHNIQUOCV		E							
DIMENSIONS ARE IN	1. INTERPRET DRAWING PER ASME Y14.5 2. REMOVE ALL SHARP EDGES, .005013 ALL EDGES APPROXIMATLEY R.02 FOR S	. FOR MACHINED PARTS. ROUND		INSTITUTE OF TECHNOLOGY SETTS INSTITUTE OF TECHNOLOGY	CUS	STOM C	ABLE S	SPE	CIFICATION	V25A-	108	
TOLERANCES: .XX ±	3. DO NOT SCALE FROM DRAWING. 4. ALL MACHINING FLUIDS MUST BE FUL	Y SYNTHETIC, FULLY WATER	SYSTEM	SUB-SYSTEM		11.71.55011	JUL/02/2012				REV.	_
.XXX ±	SOLUBLE AND FREE OF SULFUR, SILICON	E, AND CHLORINE.		130	DRAFTER	E. BROWN	JUL/02/2012	D	D110165	R	V	4
4.1.Q.III.4.D	MATERIAL	FINISH	NEXT ASSY		CHECKER				D 110103	J	▼ -	T
ANGULAR ± °	Material <not spec<="" td=""><td>fied> uind</td><td>ch </td><td></td><td>APPROVAL</td><td></td><td></td><td>SCALE:</td><td>1:1 PROJECTION:</td><td>₹ t SHI</td><td>IFFT 1 OF 1</td><td></td></not>	fied> uind	ch		APPROVAL			SCALE:	1:1 PROJECTION:	₹ t SHI	IFFT 1 OF 1	



25 COND. (12 TWISTED PAIR + 1 WIRE + SHIELD) 28AWG CABLE WITH 6 COPPER BRAID (SHIELD) AND 7 PEEK OVERBRAID COONER WIRE #CZ1105 +6 +7 **CONTINENTAL PART #24x3x40BC** COPPER BRAID - CONTINENTAL CORDAGE PART #24x3x40BC PEEK BRAID - PART #6759 MANUFACTURED WITH #6759 ZEUS 0.016" BLACK PEEK DRAWN MONOFILAMENT GLENAIR # 600-052 or GLENAIR #600-052 STANDARD BRAID CLAMP or 2 BAND-IT PART # A10086 (0.240" WIDE) ("BAG OF 100" #A10089) BAND-IT # A10086

* NOTE: USE WHATEVER LENGTH IS NECESSARY FOR THE INTERNAL WIRING OF THE CONNECTORS AND STRIP LENGTH TO **ACHIEVE THE CORRECT OVERALL LENGTHS.**

** NOTE: SEE THE "TICOR CONNECTOR PART NUMBER BUILDER" DCC#D1000219 FOR DETAILS ON THIS PART NUMBER.

NOTES: (UNLESS OTHERWISE SPECIFIED)

3

A. MATERIAL: a. J1 CONNECTOR SHELL - PEEK VICTREX 450GL30.

(TS0148-25C020BS1-225) OR EQUIVALENT **

TICOR #

(TS0148-25CG20BS1-225F) OR EQUIVALENT **

b. J2 CONNECTOR SHELL - GOLD OVER ELECTRO-LESS NICKEL SELECTIVELY METALIZED PEEK VICTREX 450GL30.

DB25 CONNECTOR BACKSHELL (NO EARS) FOR UHV (STAINLESS) WITH Ø0.225" i.d. PORT

DB25 FEMALE CONNECTOR (J2) FOR UHV (GOLD METALIZED PEEK)

DB25 CONNECTOR BACKSHELL (WITH EARS) FOR UHV (STAINLESS) WITH Ø0.225" i.d. PORT

c. BACKSHELLS - STAINLESS STEEL WITH VENT HOLE.

d. CONTACTS - BERYLLIUM COPPER ALLOY C17300,

0.000050 MIN. GOLD OVER NICKEL.

e. HARDWARE: STAINLESS STEEL, PASSIVATED. f. PEEK BRAID - PEEK VICTREX GRADE TDS-450CA30 CARBON LOADED - SUPPLIED BY LIGO.

CABLE 25 COND. 28 AWG, (65 STRD 46 AWG) WITH PFA INSULATION COONER WIRE #CZ1105. 12 TWISTED PAIRS (4 TO 5 TWISTS PER INCH) + 1 WIRE. OVERALL 40AWG COPPER BRAID 50% COVERAGE - SUPPLIED BY LIGO.

OVERALL PEEK BRAID MIN. 50% COVERAGE. OVERALL CABLE O.D. WILL BE 0.240 IN.

C. CONNECTORS WILL BE SUPPLIED WITH HARDWARE. SCREWS SHOULD BE THE PROPER LENGTH FOR MATING.

 8. ALL PARTS SHALL BE MANUFACTURED IN ACCORDANCE WITH LIGO SPECIFICATION E0900364. 9. ALL HELT-COIL HOLES TO BE PREPARED ACCORDING TO EMHART HELI-COIL PRODUCT CATALOG, HC2000, REV 4

CENTERED ON BOTH SIDES OF THE HOLE.

GLENAIR

CLAMPING **BANDS #600-052** (BAND-IT #A10086)

COPPER BRAID 1 CONDUCTOR (SHIELD) 25 CONDUCTOR 28 AWG 12 TWISTED PAIR + 1 WIRE

PEEK OVERBRAID