



LIGO Laboratory / LIGO Scientific Collaboration

LIGO-T1500404-v1

LIGO

July 25, 2015

Test Procedure for EOM/AOM Driver

Koji Arai and Daniel Sigg

Distribution of this document:
LIGO Scientific Collaboration

This is an internal working note
of the LIGO Laboratory.

California Institute of Technology
LIGO Project – MS 100-36
1200 E. California Blvd.
Pasadena, CA 91125
Phone (626) 395-2129
Fax (626) 304-9834
E-mail: info@ligo.caltech.edu

Massachusetts Institute of Technology
LIGO Project – NW22-295
185 Albany St
Cambridge, MA 02139
Phone (617) 253-4824
Fax (617) 253-7014
E-mail: info@ligo.mit.edu

LIGO Hanford Observatory
P.O. Box 159
Richland WA 99352
Phone 509-372-8106
Fax 509-372-8137

LIGO Livingston Observatory
P.O. Box 940
Livingston, LA 70754
Phone 225-686-3100
Fax 225-686-7189

<http://www.ligo.caltech.edu/>

1 Introduction

The following Test Procedure describes the test of proper operation of the EOM/AOM Drive. **The RF output port should always be properly terminated with 50 Ohms.** Further information can be found on the [wiki page](#).

S/N _____ Tester _____ Date _____

2 Test Equipment

- Voltmeter
- Oscilloscope
- Stanford Research SR785 analyzer
- RF Network Analyzer (Agilent AG4395A or equivalent)
- RF Spectrum Analyzer (Agilent AG4395A is OK)
- RF function generator (Tektronix AFG3101, or equivalent)
- AF function generator (Stanford Research DS335/340/345 or similar)
- RF Power Meter (Agilent E4418A, Gigatronics 3410A, or similar)
- D1500091 EtherCAT EOM Driver Test Rig with DB25 Male/Male cable
- DB9 breakout board
- Two dual bench power supply (for +/-24Volts, +/- 18 Volts)
- Single bench power supply (for +5V)
- Diagrams / Board Schematics, LIGO [D0900760](#), [D0900761](#), [D0900847](#), [D0900848](#)

3 Tests

The EOM/AOM Driver uses the RF AM Stabilization Power (D0900848, rev B) with the RF AM Stabilization Controller (D0900761, rev B) and RF AM Stabilization Servo (D0900847, rev B).

Power Supply Test

1) Verify the proper current draw with the output switch off:

Turn off the excitation (left) and output (right) switches. Using a bench DC supply apply +/-24Volts and +/- 18 Volts to P1 of the *RF AM Stabilization Power (D0900848)* at the *front panel*. Measure the current draw of the box. **Confirm the RF output port I terminated.** Turn on the output switch and measure the current draw of the box.

Supply Voltage	With RF output SW “OFF”		With RF output SW “ON”	
	Current	Nominal	Current	Nominal
+24V	_____	0.08A	_____	0.53A
-24V	_____	0.07A	_____	0.07A
+18V	_____	0.29A	_____	0.21A
-18V	_____	0.24A	_____	0.26A

2) Verify the internal supply voltages:

On the 2nd PCB of the power board: check the voltage between TP 1 (GND) and the following test points.

Test Point	Nominal V	Measurement	Test Point	Nominal V	Measurement
TP13	-5V	_____	TP2	+24V	_____
TP12	-15V	_____	TP3	+18V	_____
TP11	-21V	_____	TP9	+10V	_____
TP10	-10V	_____	TP17	+24V	_____
TP5	-18V	_____	TP14	+21V	_____
TP6	-24V	_____	TP15	+15V	_____
			TP16	+5V	_____

3) Verify supply OK logic:

Turn off the external power supply. Disconnect the power supply connector from the front panel. Remove the front panel. Remove the top two board from the inter PCB connector. Expose the D0900848 board. Make sure two boards have proper electrical isolation from the chassis or other conductive material. Connect and turn on the power supply again. If above TP 5, 6, 2, and 3 are correct then pin 5 on U1 (connected to R11) and U4 (connected to R23) should be

Logic high ~3Volts. Confirm._____

4) Verify the relays for the power supply sequencing:

Turn off +/-24 V. Confirm Pin 5 of K1 and K2 are not energized to +/-18V.

Confirm _____

Turn on +/-24 V again. Confirm Pin 5 of K1 and K2 are now energized to +/-18V.

Confirm _____

5) Verify noise levels of the internal power supply voltages:

Turn off +/-24 V. Disconnect the power supply from the front panel connector. Restore the two boards and the front panel.

The noise on the following test points should be measured with a SR785 using an rms power spectrum.

TP13 noise _____ less than 15 nVrms/sqrt Hz at 140 Hz

TP12 noise _____ less than 30 nVrms/sqrt Hz at 140 Hz

TP11 noise _____ less than 40 nVrms/sqrt Hz at 140 Hz

TP10 noise _____ less than 20 nVrms/sqrt Hz at 140 Hz.

TP9 noise _____ less than 20 nVrms/sqrt Hz at 140 Hz.

TP14 noise _____ less than 30 nVrms/sqrt Hz at 140 Hz.

TP15 noise _____ less than 20 nVrms/sqrt Hz at 140 Hz.

TP16 noise _____ less than 15 nVrms/sqrt Hz at 140 Hz.

RF Test**6) Test output switch functionality:**

Make sure the output switch is off. Connect the RF function generator to the input connector. Provide 9.1MHz signal at the power level of 10dBm. Connect the RF power monitor to the output connector. Set the output adjustment switches to have 13dBm output (i.e. 12dBm & 1.0dBm)

Confirm there is no output at this point.

Confirm _____

Turn on the output switch. Confirm there is the RF output about 13dBm.

Confirm _____

7) Check stabilization servo stability:

Turn off the RF output switch and connect the output to the oscilloscope with a 50 Ohm terminator. Turn the RF output again. Confirm the waveform is sinusoidal and there is no obvious amplitude oscillation.

Confirm _____

Observe the TP1 signal with the oscilloscope and SR785. Confirm there is no oscillation.

Confirm _____

Observe the CTRL signal with the oscilloscope and SR785. Confirm there is no oscillation.

Confirm _____

8) Check exc port functionality:

Observe the TP1 signal with the oscilloscope. Connect 1kHz 2Vpk signal from the audio frequency function generator to the exc port.

Turn off the exc switch and confirm there is no obvious 1kHz component on the TP1 signal.

Confirm _____

Turn on the exc switch and confirm there is visible 1kHz component on the TP1 signal.

Confirm _____

9) Openloop transfer function measurement:

Connect TP1 and TP2 to ChA and ChB of the network analyzer, respectively. Connect the signal generator output of the network analyzer to Exc. Turn on the exc switch.

Sweep the signal of 0dBm from 1kHz to 2MHz (801 line).

Measure the transfer function and check the frequency where the magnitude is 0dB (i.e. UGF: unity gain frequency). Check the phase at the UGF (i.e. P.M.: phase margin).

Confirm the UGF is above 100kHz and the phase is more than 30deg.

Unity Gain Frequency _____ (**>100kHz nom**)

Phase Margin _____ (**>40deg nom**)

10) Output calibration:

Turn off the output switch. Connect the output to the power meter. Turn on the output switch again. The power adjustment is set to be 13dBm (12dBm+1dBm). Rotate the trimmer R6 on the front panel until the output becomes 13.00dBm+/-0.01dBm.

Confirm _____

11) Check the output power with other output values:

Set the power adjustment to 4dBm+0dBm. Measure and record the RF output power with the power meter, while sweeping the output from 4+0dBm to 26+0dBm.

Setting (dBm)	Measurement (dBm)	Nominal (dBm)
4.0	_____	4.0+/-0.1
6.0	_____	6.0+/-0.1
8.0	_____	8.0+/-0.1
10.0	_____	10.0+/-0.1
12.0	_____	12.0+/-0.1
14.0	_____	14.0+/-0.1
16.0	_____	16.0+/-0.1
18.0	_____	18.0+/-0.1
20.0	_____	20.0+/-0.1
22.0	_____	22.0+/-0.1
24.0	_____	>23.5
26.0	_____	>24.0

Set the coarse power adjustment to 4dBm, 16dBm, 22dBm and scan the fine adjustment from 0dBm to 2.2dBm. For each scan, measure and record the RF output power with the power meter.

4dBm			16dBm			26dBm		
Setting (dBm)	Measurement (dBm)	Nominal (dBm)	Setting (dBm)	Measurement (dBm)	Nominal (dBm)	Setting (dBm)	Measurement (dBm)	Nominal (dBm)
+0.0	_____	4.0+/-0.1	+0.0	_____	16.0+/-0.1	+0.0	_____	>24.5
+0.2	_____	4.2+/-0.1	+0.2	_____	16.2+/-0.1	+0.2	_____	>24.5
+0.4	_____	4.4+/-0.1	+0.4	_____	16.4+/-0.1	+0.4	_____	>24.5
+0.6	_____	4.6+/-0.1	+0.6	_____	16.6+/-0.1	+0.6	_____	>24.5
+0.8	_____	4.8+/-0.1	+0.8	_____	16.8+/-0.1	+0.8	_____	>24.5
+1.0	_____	5.0+/-0.1	+1.0	_____	17.0+/-0.1	+1.0	_____	>24.5
+1.2	_____	5.2+/-0.1	+1.2	_____	17.2+/-0.1	+1.2	_____	>24.5
+1.4	_____	5.4+/-0.1	+1.4	_____	17.4+/-0.1	+1.4	_____	~4.0
+1.6	_____	5.6+/-0.1	+1.6	_____	17.6+/-0.1	+1.6	_____	~4.0
+1.8	_____	5.8+/-0.1	+1.8	_____	17.8+/-0.1	+1.8	_____	~4.0
+2.0	_____	6.0+/-0.1	+2.0	_____	18.0+/-0.1	+2.0	_____	~4.0
+2.2	_____	6.2+/-0.1	+2.2	_____	18.2+/-0.1	+2.2	_____	~4.0

Stabilization Test**12) Check monitor values:**

Keep the RF output connected to the power meter. Set the output level at 13dBm. Check voltage values of the monitor channels with the voltage meter.

Note: BIAS1 and MON1 are located at the right half of the top monitor BNC connectors. BIAS2 and MON2 are at the left half.

Channel	Measurement	Nominal
CTRL	_____	~3V
BIAS1	_____	1.7~1.8V
MON1	_____	< +/-15mV
BIAS2	_____	1.7~1.8V
MON2	_____	< +/-15mV

13) Closed loop transfer function of the internal loop for the 2nd RFAM detector

Keep 13dBm output of the unit on the power meter. Connect the signal generator of SR785 to the exc port and inject 5Vpk random noise. Set the frequency range upto 25Hz and 800lines. Turn on the exc switch. Measure the transfer function from MON1 to MON2.

Confirm the transfer function is 0 dB +/- 0.5 dB above 10Hz.

Confirm _____

14) Power Spectral Density

Keep 13dBm output of the unit on the power meter. Make sure that the exc switch is off and the signal generator of SR785 is disconnected from the exc port. Measure power spectral density of the following signals at the listed frequencies. Note: PSD Unit (/rtHz) should be turned **on** when reading the numbers.

Channel	Measurement [V/rtHz]			
	10Hz	100Hz	1kHz	10kHz
CTRL	_____	_____	_____	_____
BIAS1	_____	_____	_____	_____
MON1	_____ (14A)	_____ (14B)	_____ (14C)	_____ (14D)
BIAS2	_____	_____	_____	_____
MON2	_____ (14E)	_____ (14F)	_____ (14G)	_____ (14H)

15) Calibration 1

Keep 13dBm output of the unit on the power meter. Connect the signal generator of SR785 to the exc port and inject 5Vpk sinusoidal signal. Turn on the exc switch. Check the signal amplitude of the following ports at 1kHz. Note: PSD Unit (/rtHz) should be turned **off** when reading the numbers.

Channel	Measurement [mVrms]	Nominal
CTRL	_____	~3mVrms
MON1	_____ (15A)	~350mVrms
MON2	_____ (15B)	~350mVrms

16) Calibration 2

Keep the above excitation setting. Turn off the output switch and connect the output to the spectrum analyzer. Turn on the output switch again. Measure the peak height of the carrier power at 9.1MHz and sideband power at 9.099MHz and 9.101MHz.

Signal	Measurement [dBm]
Carrier (9.100MHz)	_____ (16A)
Upper Sideband (9.101MHz)	_____ (16B)
Lower Sideband (9.099MHz)	_____ (16C)

17) Calibration ~ Calculation

Copy numbers from above calibration measurements and fill the blanks. Confirm the final numbers if they fulfill the requirements

Signal Amplitudes

$$A_{CR} = \text{Sqrt}[10^{(\text{_____}_{(16A)} / 10) / 1000 * 50 * 2}] = \text{_____} [\text{Vpk}]$$

$$A_{USB} = \text{Sqrt}[10^{(\text{_____}_{(16B)} / 10) / 1000 * 50 * 2}] = \text{_____} [\text{Vpk}]$$

$$A_{LSB} = \text{Sqrt}[10^{(\text{_____}_{(16C)} / 10) / 1000 * 50 * 2}] = \text{_____} [\text{Vpk}]$$

Modulation Depth

$$m_{AM} = (A_{USB} + A_{LSB}) / A_{CR} = \text{_____} [\text{RIN}]$$

Calibrations of the monitor channels

$$\text{cal}_{MON1} = \text{_____}_{(15A)} * \text{Sqrt}(2) / m_{AM} = \text{_____} [\text{Vpk/RIN}]$$

$$\text{cal}_{MON2} = \text{_____}_{(15B)} * \text{Sqrt}(2) / m_{AM} = \text{_____} [\text{Vpk/RIN}]$$

Calibrated power spectral densities

MON1 (In-loop)

10Hz: _____^(14A) / cal_{MON1} = _____ [RIN / rtHz] less than 1.0×10^{-8} /rtHz (160dBc/Hz)

100Hz: _____^(14B) / cal_{MON1} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

1kHz: _____^(14C) / cal_{MON1} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

10kHz: _____^(14D) / cal_{MON1} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

MON2 (Out-of-loop)

10Hz: _____^(14E) / cal_{MON2} = _____ [RIN / rtHz] less than 1.0×10^{-8} /rtHz (160dBc/Hz)

100Hz: _____^(14F) / cal_{MON2} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

1kHz: _____^(14G) / cal_{MON2} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

10kHz: _____^(14H) / cal_{MON2} = _____ [RIN / rtHz] less than 3.2×10^{-9} /rtHz (170dBc/Hz)

DAQ channel Test**18) Check functionality of the DAQ port**

Measure the transfer functions between the following ports with SR785 with no excitation injected. The two pins of the DAQ port signals should be measured differentially: Use “A-B” input mode of the SR785. Connect two signal clips to the SR785 inputs and clip the signals with the positive clips.

Transfer Function		Measurement	Nominal Gains
From (BNC Port)	To (DAQ DSUB Port)		
MON1	Pin1(+) and Pin6 (-)	_____	20?
CTRL	Pin2(+) and Pin7 (-)	_____	2 (f<0.1Hz), 40 (f>100Hz)
MON2	Pin3(+) and Pin8 (-)	_____	20?
BIAS2	Pin4(+) and Pin9 (-)	_____	2

Test with Remote Control RIG**19) Check indicators of EtherCAT EOM Driver Test Rig**

Continue to use the 13dBm output setting.

Turn off the exc and output switches. Make sure the Int/Ext switch on the main unit is at “Int”. Connect the test rig to the control connector with a DB25 M/M cable.

Make sure the Latch switch on the test rig is at “Dis”.

Toggle the Int/Ext switch on the main unit. Confirm the indicator “Int” on the test rig correctly shows the state. **Confirm** _____

Turn the Int/Ext switch on the main unit to “Ext”.

Toggle the RF output switch on the main unit. Confirm the indicator “RF” on the test rig correctly shows the state. **Confirm** _____

Make sure the exc switch is off. Toggle the Exc Enable on and off. Confirm the indicator “Exc Readback” on the test rig correctly shows the state. **Confirm** _____

20) Power setting test

Turn off the RF output switch on the main unit. Connect the power meter to the RF output port. Then turn on the RF output switch. Turn off (up) all the value switches.

Measure the RF output power at this point and confirm it is still about 13dBm

Confirm _____

Turn the Latch switch on the test rig to the “En (Enable)” position.

Make sure the RF output power is now about 4dBm

Confirm _____

Measure the RF output while turning on the switches one by one. (i.e. Only one switch is on at a time). Note: The panel marking (+1dB, +2dB, ...) is obsolete and they should be scaled by a factor of 0.2. (i.e. +1dB = +0.2dB, +2dB = +0.4dB, and so on)

Switch	Measurement	Nominal [dBm]
+1dB on	_____	4.2+/-0.1
+2dB on	_____	4.4+/-0.1
+4dB on	_____	4.8+/-0.1
+8dB on	_____	5.6+/-0.1
+16dB on	_____	7.2+/-0.1
+32dB on	_____	10.4+/-0.1
+64dB on	_____	16.8+/-0.1
+128dB on	_____	~4.0

21) Test on the signals via the test rig

Turn on the +1, +4, +8, and +32dB switches on the test rig. This should yield the output of 13dBm. Measure the transfer functions between the following ports with SR785.

Transfer Function		Measurement	Nominal Gains
From (BNC Port)	To (TEST RIG Port)		
BIAS1	BIAS1	_____	1
BIAS2	BIAS2	_____	1
CTRL	CTRL	_____	Bandpass shape 20dB@15Hz 0deg