



Advanced LIGO Photon Calibrators

Sudarshan Karki
University of Oregon

GW Metrology Workshop
NIST, Boulder, CO
March 15, 2019








Introduce fiducial displacements


$$x(f) = \frac{2 \cos(\theta)}{c} P(f) \frac{1}{M(2\pi f)^2} \mathcal{R} \mathcal{G}(f)$$

Local Elastic Deformation + Rotation




Bulk Elastic Deformation






Schematic Layout of Photon calibrator




Bird's eye view of the layout of Photon calibrator as installed.




Pcal persicope structure


$$x(f) = \frac{2 \cos(\theta)}{c} \boxed{P(f)} \frac{1}{M(2\pi f)^2} \mathcal{R} \mathcal{G}(f)$$



- Gold Standard (GS) calibrated at NIST
 - » One single standard.
- Working Standard (WS) calibrated against GS.
 - » WS -> One for each detector
- Pcal power sensors (Tx and Rx) at each end station calibrated against WS.



$$\rho_R = \alpha_{RW} \times \alpha_{WG} \times \rho_G$$






GS → WS


- Annual NIST measurement has uncertainty of $\sim 0.35\%$.
- Uncertainties in relative responsivity measurements are approx. 0.10% .





WS → Tx / Rx



Parameter	Relative Uncertainty
NIST \rightarrow GS [ρ_{GS}]	0.35 %
WS/GS [α_{WG}]	0.10 %
Rx/WS [α_{RW}]	0.10 %
Overall	< 0.40 %



$$\frac{1.3\%}{2\sqrt{3}} \approx 0.37\%$$



In vacuum measurements at all 4 end stations

- » Allows us to apportion the losses between the input and output paths


0.37% → 0.10%

$$x(f) = \frac{2 \cos(\theta)}{c} P(f) \frac{1}{M(2\pi f)^2} \mathcal{R} \mathcal{G}(f)$$


Parameter	Rel. Uncertainty
Mass (M)	0.005 %
Cos θ	0.07 %

$$x(f) = \frac{2 \cos(\theta)}{c} P(f) \frac{1}{M(2\pi f)^2} \mathcal{R} \mathcal{G}(f)$$

- Unintended Rotational effect
 - » Poor localization of the beams
 - » Power imbalance between the beams





$$\mathcal{R}(a, b) = \left[1 + \frac{M}{I} \vec{a} \cdot \vec{b} \right]$$



$\sim 0.20 \%$


Parameter	Relative Uncertainty
NIST \rightarrow GS [ρ_{GS}]	0.35 %
WS/GS [α_{WG}]	0.10 %
Rx/WS [α_{RW}]	0.10 %
Optical efficiency [\mathcal{E}_T]	0.10 %
Angle of incidence [$\cos \theta$]	0.07 %
Mass of test mass [M]	0.005 %
Rotation [\mathcal{R}]	0.20 %
Overall	< 0.50 %

This is the accuracy of calibration on the displacement fiducials.




Frequency Response of Working Standard


Total Harmonic Distortion (14 harmonics)



Error less than 0.01% up to 7 kHz



Butterfly Mode
5953 Hz





Drumhead Mode
8151 Hz

ETM motion deviates from their rigid body approximation due to the excitation of the natural modes by applied forces


$$x(f) = \frac{2 \cos(\theta)}{c} P(f) \frac{1}{M(2\pi f)^2} \mathcal{R} \boxed{\mathcal{G}(f)}$$

LIGO CALIBRATION AT HIGHER FREQUENCIES


FEA (COMSOL) Results



Front face of the test mass



Photon calibrators, that rely on absolute laser power calibration from NIST, are providing fiducial displacements with sub-1% accuracy.






EXTRA SLIDES





$$\vec{a} = \frac{\beta \vec{a}_1 + \vec{a}_2}{\beta + 1}$$





- Principle of Operation
- Hardware Overview
- Laser Power and Displacement Calibration
 - » Uncertainty
- Frequency Dependence
- Summary

Experimental confirmation of the results estimated from FEA.




Enable better calibration at high frequencies.






Improvement in calibrated signal after applying the
time-varying calibration parameters




Unusually large variations in NIST calibration




Comparison of Global Power Standard.

- The larger variations in NIST measurement was due to the light reflecting back from the nearby aperture into the power sensor.
 - » Confirmed this through an experiment conducted at NIST during the visit.
 - » Will start new measurements of Gold Standard and collect at least two data points before O3.
- Calibration provided by NIST are traceable to fundamental units.
 - » They are the only institute that calibrate at 1 W level.
 - » Rest of the countries extrapolate from microwatt level

Config.	Pcal Beam Position			Beam Config.
	Beam	Target (mm)	Actual (mm)	
OP	Upper	[0, 111.6]	[0.8, 112.2]	
	Lower	[0, -111.6]	[-0.8, -111.8]	
P14	Upper	[0, 127.6]	[1.2, 126.1]	
	Lower	[0, -127.6]	[-0.5, -125.7]	
M8	Upper	[0, 103.6]	[2.5, 103.2]	
	Lower	[0, -103.6]	[-1.3, -103.0]	

Cahillane et al., Phys Rev D 96, 102001







Observing Run 1

Observing Run 2


Path	Tx \rightarrow ETM	ETM refl.	ETM \rightarrow Rx	Total (Tx \rightarrow Rx)
Inner Beam	0.9947	1.0007	0.9957	0.9911
Outer Beam	0.9920	1.0009	0.9933	0.9862
Average	0.9933	1.0008	0.9945	0.9887
Total optical loss				1.13 %
Fraction of optical Loss on Input side				0.56
Fraction of optical Loss on Output side				0.44

With 55% to
45% loss on
two sides and





WS/GS \rightarrow 0.03%



RX/WS → 0.05%